

# Defibtech Lifeline VIEW Semi-Automatic Defibrillator

## TECHNICAL SPECIFICATIONS†

### DEFIBRILLATOR

#### TYPE

Semi-automatic external defibrillator

#### MODEL

DDU-2300

#### WAVEFORM

Biphasic Truncated Exponential  
(Impedance compensated)

#### ENERGY

Adult: 150 Joules  
Child / Infant: 50 Joules  
(Nominal into 50 Ohm load)

#### CONTROLS

Lighted On/Off button  
Lighted Shock button

#### CHARGE TIME\*

Less than 4 seconds  
(from shock advised)

#### DISPLAY

High-resolution color LCD

#### VIDEO PROMPTS

Full motion video  
On-screen text prompts

#### CPR COACHING

Video and voice coaching  
On-demand video help

#### VOICE PROMPTS

Extensive voice prompts guide  
user through operation of the unit

#### RESCUE PROTOCOL

AHA/ERC (default);  
supports protocol updates by  
the user (password protected)

\*Typical, with new battery at 25°C

### DEFIBRILLATION / MONITORING PADS

#### MODEL

Adult: DDP-2001  
Child / Infant: DDP-2002

#### SURFACE AREA\*\*

Adult: 12 inches<sup>2</sup> (77 cm<sup>2</sup>)  
Child / Infant:  
7.75 inches<sup>2</sup> (50 cm<sup>2</sup>)

#### TYPE

Pre-connected, single-use,  
non-polarized, disposable,  
self-adhesive electrodes with  
cable and connector

\*\*Nominal, each pad

### EVENT DOCUMENTATION

#### INTERNAL EVENT RECORD

Critical ECG segments and rescue  
event parameters are recorded  
(greater than 60 minutes) and  
can be downloaded to a removable  
data card

#### PC-BASED EVENT REVIEW

ECG with event tag display, and  
audio playback when available

#### REMOVABLE STORAGE

(optional) Up to 30 hours of ECG  
and event data storage (no audio  
option) or up to 3 hours of audio  
(audio option). ECG and event  
storage on a removable data card.  
Actual length of storage is  
dependent on card capacity.

#### USB PORT

Event download and  
maintenance operations

### PATIENT ANALYSIS SYSTEM

#### PATIENT ANALYSIS

Automatically evaluates patient  
impedance for proper pad contact.  
Monitors signal quality and analyzes  
patient ECG for shockable/  
non-shockable rhythms.

#### SENSITIVITY/SPECIFICITY

Meets or exceeds IEC-60601-2-4  
requirements; meets AAMI DF80  
requirements and AHA  
recommendations

### BATTERY PACK

#### MODEL

DBP-2003 (standard),  
DBP-2013 (aviation; TSO C-142a)

#### POWER

12V, 2800 mAh

#### TYPE

Lithium/Manganese Dioxide  
Disposable, recyclable,  
non-rechargeable

#### CAPACITY\*

125 shocks or 8 hours  
continuous operation

#### STANDBY LIFE\*

4 years

#### LOW BATTERY INDICATORS

Visible  
Audible

\*Typical, with new battery at 25°C

### SELF TESTS

#### AUTOMATIC

Automatic daily, weekly, monthly  
and quarterly circuitry tests

#### BATTERY INSERTION

System integrity test on  
battery insertion

#### PAD PRESENCE

Pads preconnected tested daily

#### USER-INITIATED

Unit and battery pack system  
test initiated by the user

#### STATUS INDICATION

Visual and audible indication  
of unit status

#### STATUS SCREEN

Unit self-test results  
Pads and battery information  
(status and expiration)

### ENVIRONMENTAL

#### TEMPERATURE

Operating: 0 to 50°C (32 to 122°F)  
One Hour Operating Temperature  
Limit (extreme cold):  
-20°C (-4°F)\*\*\*  
Standby: 0 to 50°C (32 to 122°F)

#### RELATIVE HUMIDITY

Operating / Standby: 5%-95%  
(non-condensing)

#### ALTITUDE

-500 to 15,000 ft (-150 to 4500 m)  
per MIL-STD-810F 500.4 Procedure II

#### VIBRATION

Ground (MIL-STD-810F 514.5  
Category 20)

Helicopter (RTCA/DO-160D,  
Section 8.8.2, Cat R, Zone 2,  
Curve G)

Jet Aircraft (RTCA/DO-160D  
Section 8, Cat H, Zone 2,  
Curves B & R)

#### SHOCK / DROP ABUSE TOLERANCE

MIL-STD-810F 516.5 Procedure IV  
48 inches (1.2 meters), any edge,  
corner, or surface, in standby mode

#### SEALING / WATER RESISTANCE

IEC 60529 class IP55;  
Dust Protected, Protected against  
water jets (Battery pack installed)

#### ESD

IEC 61000-4-2: (Open air up to  
15kV or direct contact up to 8kV)

#### EMC (Emission)

CISPR 11 Group 1 Level B  
and FCC Part 15

#### EMC (Immunity)

IEC 61000-4-3 and IEC 61000-4-8

\*\*\*From room temperature to  
temperature extreme, one hour  
duration, updated specification for  
DDU-2000 Series AEDs running  
software revision 2.4 or above

### PHYSICAL

#### SIZE

7.3 x 9.5 x 2.3 Inches  
(18.5 x 24 x 5.8 cm)

#### WEIGHT (with battery)

Less than 3 lbs (1.4 kg)

## BRIEF SUMMARY OF INDICATIONS, CONTRAINDICATIONS AND OTHER IMPORTANT SAFETY INFORMATION

### When should the Defibtech Automated External Defibrillator (AED) be used - what are its indications?

Lifeline/ReviveR VIEW DDU-2300 and Lifeline/ReviveR ECG DDU-2450 Automated External Defibrillators (AEDs) are indicated for use on victims of sudden cardiac arrest (SCA) who are:

- Unconscious and unresponsive
- Not breathing or not breathing normally

Lifeline/ReviveR VIEW DDU-2300 and Lifeline/ReviveR ECG DDU-2450 AEDs may be used with Defibtech adult defibrillation pads (model number DDP-2001). For patients under 8 years old, or weighing less than 55 lbs (25 kg), use Defibtech child/infant defibrillation pads (model number DDP-2002), if available.

Lifeline/ReviveR VIEW DDU-2300 and Lifeline/ReviveR ECG DDU-2450 Automated External Defibrillators (AEDs) should not be used if the victim is responsive or conscious.

### What other information is important about using the AED?

Do not delay therapy to determine exact age or weight. If pediatric pads are not available, apply adult pads in the position as shown for a child/infant and use the AED.

### What are the potential adverse health effects of using an AED?

The potential adverse effects (e.g., complications) associated with use of an automated external defibrillator include, but are not limited to, the following:

- Failure to identify shockable arrhythmia.
- Failure to deliver a defibrillation shock in the presence of VF or pulseless VT, which may result in death or permanent injury.
- Inappropriate energy, which could cause failed defibrillation or post-shock dysfunction.
- Myocardial damage.
- Fire hazard in the presence of high oxygen concentration or flammable anesthetic agents.
- Incorrectly shocking a pulse sustaining rhythm and inducing VF or cardiac arrest.
- Bystander shock from patient contact during defibrillation shock.
- Interaction with pacemakers.
- Skin burns around the defibrillation pads placement area.
- Allergic dermatitis due to sensitivity to the materials used in the defibrillation pads construction.
- Minor skin rash.

### What are some of the relevant warnings related to the AED?

- Hazardous electrical output. This equipment is for use only by qualified personnel.
- Possible fire or explosion. Do not use in the presence of flammable gases or anesthetics. Use care when operating this device close to oxygen sources (such as bag-valve-mask devices or ventilator tubing). Turn off gas source or move source away from patient during defibrillation, if necessary.
- The DDU-2000 Series AED has not been evaluated or approved for use in hazardous locations as defined in the National Electric Code standard. In compliance with IEC classification, the DDU-2000 Series AED is not to be used in the presence of flammable substance/air mixtures.
- Improper maintenance can cause the DDU-2000 Series AED not to function. Maintain the DDU-2000 Series AED only as described in the User Manual and Operating Guide. The AED contains no user-serviceable parts — do not take the unit apart.
- Do not open sealed pads package until pads are to be used. The packaging should be opened only immediately prior to use, otherwise the pads may dry out and become non-functional.
- Do not touch the patient during defibrillation. Defibrillation current can cause operator or bystander injury.
- The defibrillation pads are intended for one-time use only and must be discarded after use. Reuse can lead to potential cross infection, improper performance of the device, inadequate delivery of therapy, and/or injury to the patient or operator.
- CPR during analysis can cause incorrect or delayed diagnosis by the patient analysis system.
- User-initiated and automatic self-tests are designed to assess the DDU-2000 Series AED's readiness for use. However, no degree of testing can assure performance or detect abuse, damage, or a defect that occurred after the most recent test is completed.
- Even if defibrillation occurs, the sudden cardiac arrest event may not result in survival.

### What are some of the relevant cautions related to the AED?

- Follow all battery pack labeling instructions. Do not install battery packs after the expiration date.
- Follow all defibrillation pad label instructions. Use defibrillation pads prior to their expiration date.
- Use and store the DDU-2000 Series AED only within the range of environmental conditions specified in the technical specifications.

**Caution:** Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Please refer to the Operating Guide provided with your AED for user instructions, complete list of warnings and cautions, operator training requirements, summary of primary clinical studies, technical specifications, and other important information. The Operating Guide, for concise guidance on set-up, use, maintenance and technical specifications, and User Manual, for comprehensive training on set-up, use and maintenance; and source for complete technical specifications, are also available at [www.defibtech.com/support](http://www.defibtech.com/support).



Defibtech, LLC • Guilford, CT 06437 USA • 1-203-453-4507 • 1-866-DEFIB-4U (1-866-333-4248)  
[www.defibtech.com](http://www.defibtech.com)

ELECTRONIC DISTRIBUTION

DAC-A2703EN-BL